

## **Product Specification**

## Ammonium Hydroxide, 5% (v/v)

Lot Number: SAMPLE Prod	duct Number: 628		Manufacture Date: N/A Expiration Date: N/A
Name	CAS#	Grade	
Water	7732-18-5	ACS/ASTM/USP/EP	
Ammonium Hydroxide	1336-21-6	ACS	
Test	Specification	Result	NIST SRM#
Appearance	Colorless liquid	N/A	
Assay (vs. Sulfuric Acid/Methyl Red)	4.9 <b>-</b> 5.1 % (v/v)	N/A	723
Specification	Reference		
Ammonium hydroxide, 5%	TAPPI (T 618 cm-84)		
Ammonium Hydroxide, 1+19	APHA (3500-Mg D)		
Ammonium Hydroxide (1+19)	ASTM (D 2187 H)		
Volumetric glassware complies with Class A tole recalibrated regularly in accordance with ASTM traceable to the NIST national mass standard. T thermometer traceable to NIST standards. All pr methods. Batch records document raw material t <b>Part Number</b>	E 542 and NIST Procedure NBSIR 74-46 hermometers and temperature probes ar oducts are prepared according to master	31. Balances are calibrated regu- re calibrated before first use and documents that assure manufa istory for each lot manufacture	larly with weights certified l recalibrated regularly with a acture according to validated
		24 months	
628-1 628-16	4 L natural poly	24 months 24 months	
628-32	500 mL natural poly 1 L natural poly	24 months 24 months	
628-5	20 L Cubitainer®	24 months 24 months	
628-55		24 months	
040 00	55 gal poly drum	24 III0110115	

Recommended Storage: 15°C - 30°C (59°F - 86°F)

This document is designed to comply with ISO Guide 31 "Reference Materials --Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.