

Product Specification

Ammonium Thiocyanate, 30% (w/v)

Lot Number: SAMPLE

Product Number: 698

Manufacture Date: N/A

Expiration Date: N/A

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Ammonium Thiocyanate	1762-95-4	ACS

Test	Specification	Result
Appearance	Colorless to light pink liquid	N/A
Suitability for Use	Red color with Iron (Fe)	N/A

Specification	Reference
Ammonium Thiocyanate Solution (300 g/L)	ASTM (E 200)
Ammonium Thiocyanate Reagent Solution	ACS (N/A)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
698-1	4 L natural poly	18 months
698-16	500 mL natural poly	18 months
698-32	1 L natural poly	18 months
698-4	120 mL natural poly	18 months
698-5	20 L Cubitainer®	18 months
698-8	250 mL natural poly	18 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.