

# Product Specification

## Arsenic Standard, 0.5 ppm As in 2% HNO<sub>3</sub>

**Lot Number:** SAMPLE

**Product Number:** R0797005

**Manufacture Date:** N/A

**Expiration Date:** N/A

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Nitric Acid	7697-37-2	Trace Metals
Arsenic	7440-38-2	High Purity

Test	Specification	Result	NIST SRM#
Appearance	Colorless liquid	N/A	
Arsenic (As)	0.450-0.550 ppm	N/A	3103

This standard is guaranteed to be stable and accurate provided the product is kept tightly capped and stored under normal laboratory conditions. Balances are calibrated using NIST traceable weights whose verification of maintenance and recalibration is documented per in-house Standard Operating Procedures. Class A glassware is also calibrated and routinely rechecked per in-house Standard Operating Procedures. Trace metal analyzed acids and Trace Metals Analyzed Water are used in the manufacture of this product. Triple cleaned containers are used in the manufacture of this product.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
R0797005-500N	500 mL natural poly	12 months

**Recommended Storage:** 15°C - 30°C (59°F - 86°F)

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.