Arlington, TX 76012
Pocomoke City, MD 21851
Batesville, IN 47006
http://www.riccachemical.com
1-888-GO-RICCA
customerservice@riccachemical.com

## **Product Specification**

Buffer Solution, Giordano, pH 6.4 (M/15), for Wright Staining

Lot Number: SAMPLE Product Number: 1450

Manufacture Date: N/A Expiration Date: N/A

For In Vitro Diagnostic Use. This product is suitable for use in Wright, Wright-Giemsa, and Giemsa staining. A recommended procedure for Wright staining is available in RICCA CHEMICAL COMPANY Technical Reference Document TRD09. A recommended procedure for Wright-Giemsa staining is available in TRD10. A recommended procedure for Giemsa staining is available in TRD11. This product may also be suitable for use in other methods.

| Name                           | CAS# Grade |                 |
|--------------------------------|------------|-----------------|
| Water                          | 7732-18-5  | ACS/ASTM/USP/EP |
| Potassium Dihydrogen Phosphate | 7778-77-0  | ACS             |
| Sodium Phosphate Dibasic       | 7558-79-4  | ACS             |

| Test                                  | Specification    | Result | NIST SRM#                     |
|---------------------------------------|------------------|--------|-------------------------------|
| Appearance                            | Colorless liquid | N/A    |                               |
| pH at 25°C (Method: SQCP027, SQCP033) | 6.35-6.45        | N/A    | 185i,<br>186-I-g,<br>186-II-g |

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

| Part Number | Size / Package Type | Shelf Life (Unopened Container) |
|-------------|---------------------|---------------------------------|
| 1450-1      | 4 L natural poly    | 24 months                       |
| 1450-16     | 500 mL natural poly | 24 months                       |
| 1450-32     | 1 L natural poly    | 24 months                       |

Recommended Storage: 15°C - 30°C (59°F - 86°F)

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.

Version: 1.3 Lot Number: SAMPLE Product Number: 1450 Page 1 of 1