

Product Specification

Citric/Phosphate Buffer, pH 6.0

Lot Number: SAMPLE

Product Number: R2114800

Manufacture Date: N/A

Recommended Retest Date: N/A

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Sodium Phosphate Dibasic	7558-79-4	ACS
Citric Acid Monohydrate	5949-29-1	ACS

Test	Specification	Result	NIST SRM#
Appearance	clear, colorless	N/A	
Density at 25°C	1.0-3.8 °Be	N/A	
pH at 25°C (Method: SQCP027, SQCP033)	5.95-6.05	N/A	185i, 186-I-g, 186-II-g
Specific Gravity at 25°C	1.007-1.027	N/A	

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
R2114800-10F	10 L Cubitainer®	12 months
R2114800-4A	4 L natural poly	12 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.