

Product Specification

Potassium Chloride, 3 Molar, Filling Solution for ROSS pH Combination and Reference Electrodes

Lot Number: SAMPLE

Product Number: 5899

Manufacture Date: N/A

Expiration Date: N/A

All lots of this product are certified suitable for refilling ROSS® Type pH Combination Electrodes and ROSS® Type pH Reference Electrodes. This product can also be used for other applications where the concentration and matrix are appropriate.

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Potassium Chloride	7447-40-7	ACS

Test	Specification	Result
Appearance	Colorless liquid	N/A
Assay (by Refractive Index)	2.93-3.07 Molar	N/A

Specification	Reference
---------------	-----------

Potassium Chloride, Aqueous (3.0 M)

ASTM (D 94 B)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
5899-1	4 L natural poly	24 months
5899-16	500 mL natural poly	24 months
5899-32	1 L natural poly	24 months
5899-4	120 mL natural poly	24 months
5899-5	20 L Cubitainer®	24 months
R5899000-500A	500 mL natural poly	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.