Arlington, TX 76012
Pocomoke City, MD 21851
Batesville, IN 47006
http://www.riccachemical.com
1-888-GO-RICCA
customerservice@riccachemical.com

## **Product Specification**

0.1 ppm Mixed Metals (Ni, Ag, Cu, Cr), 1 ppm Cd in 2% Nitric Acid

Lot Number: SAMPLE Product Number: RICP5011

Manufacture Date: N/A Expiration Date: N/A

This is a multielement blend solution that was prepared volumetrically to contain the certified values reported. The uncertainty associated with the certified values is the sum of the estimated errors due to the purity of the raw material, the volumetric preparation of the solution, and transpiration of the solution through the container wall.

The final solution concentrations are confirmed by AA, ICP, or ICP-MS.

Matrix: 2% Nitric Acid

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Nitric Acid	7697-37-2	Trace Metals

Test	Specification	Result
Appearance	Colorless liquid	N/A

Analyte	Analysis (ppm)	Solute	CAS#	Grade	NIST SRM#
Chromium (Cr)	0.09-0.11 ppm	Chromium Nitrate Nonahydrate	7789-02-8	High Purity	3112
Cadmium (Cd)	0.9-1.1 ppm	Cadmium	7440-43-9	High Purity	3108
Copper (Cu)	0.09-0.11 ppm	Copper	7440-50-8	High Purity	3114
Silver (Ag)	0.09-0.11 ppm	Silver	7440-22-4	High Purity	3151
Nickel (Ni)	0.09-0.11 ppm	Nickel	7440-02-0	High Purity	3136

This standard is guaranteed to be stable and accurate provided the product is kept tightly capped and stored under normal laboratory conditions. Balances are calibrated using NIST traceable weights whose verification of maintenance and recalibration is documented per in-house Standard Operating Procedures. Class A glassware is also calibrated and routinely rechecked per in-house Standard Operating Procedures. Trace metal analyzed acids and Trace Metals Analyzed Water are used in the manufacture of this product. Triple cleaned containers are used in the manufacture of this product.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
RICP5011-500A	500 mL natural poly	12 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

Version: 1.5 Lot Number: SAMPLE Product Number: RICP5011 Page 1 of 2

This document is designed to com Contents of Certificates and Labe	ply with ISO Guide 31 "Reference Mate ls."	rials	
This test report shall not be repro	duced, except in full, without the writte	en approval of Ricca Chemical Company.	
Version: 1.5	Lot Number: SAMPLE	Product Number: RICP5011	Page 2 of 2