

## **Product Specification**

## 2 ppm Mixed Metals (Ni, Ag, Cu, Cr) 0.1 ppm Cd in 2 % Nitric Acid

## Lot Number: SAMPLE

Product Number: RICP5020

Arlington, TX 76012 Pocomoke City, MD 21851 Batesville, IN 47006 http://www.riccachemical.com 1-888-GO-RICCA customerservice@riccachemical.com

## Manufacture Date: N/A Expiration Date: N/A

This is a multielement blend solution that was prepared volumetrically to contain the certified values reported. The uncertainty associated with the certified values is the sum of the estimated errors due to the purity of the raw material, the volumetric preparation of the solution, and transpiration of the solution through the container wall.

The final solution concentrations are confirmed by AA, ICP, or ICP-MS.

Matrix: 2% Nitric Acid

| Name        | CAS#      | Grade           |
|-------------|-----------|-----------------|
| Water       | 7732-18-5 | ACS/ASTM/USP/EP |
| Nitric Acid | 7697-37-2 | Trace Metals    |

| Test       | Specification    | Result |
|------------|------------------|--------|
| Appearance | Colorless liquid | N/A    |

| Analyte       | Analysis (ppm) | Solute                          | CAS#      | Grade       | NIST SRM# |
|---------------|----------------|---------------------------------|-----------|-------------|-----------|
| Chromium (Cr) | 1.8-2.2 ppm    | Chromium Nitrate<br>Nonahydrate | 7789-02-8 | High Purity | 3112      |
| Copper (Cu)   | 1.8-2.2 ppm    | Copper                          | 7440-50-8 | High Purity | 3114      |
| Silver (Ag)   | 1.8-2.2 ppm    | Silver                          | 7440-22-4 | High Purity | 3151      |
| Nickel (Ni)   | 1.8-2.2 ppm    | Nickel                          | 7440-02-0 | High Purity | 3136      |
| Cadmium (Cd)  | 0.09-0.11 ppm  | Cadmium                         | 7440-43-9 | High Purity | 3108      |

This standard is guaranteed to be stable and accurate provided the product is kept tightly capped and stored under normal laboratory conditions. Balances are calibrated using NIST traceable weights whose verification of maintenance and recalibration is documented per in-house Standard Operating Procedures. Class A glassware is also calibrated and routinely rechecked per in-house Standard Operating Procedures. Trace metal analyzed acids and Trace Metals Analyzed Water are used in the manufacture of this product. Triple cleaned containers are used in the manufacture of this product.

| Part Number   | Size / Package Type | Shelf Life (Unopened Container) |
|---------------|---------------------|---------------------------------|
| RICP5020-500A | 500 mL natural poly | 12 months                       |

Recommended Storage: 15°C - 30°C (59°F - 86°F)

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.