

# Product Specification

**HydroSpec® Water Standard 0.10 mg/g**  
**Standard for Coulometric Karl Fischer Titration**

Lot Number: SAMPLE

Product Number: RK410000

Manufacture Date: N/A

Recommended Retest Date: N/A

Name	CAS#	Grade
Xylene	1330-20-7	
Ethylbenzene	100-41-4	
Water	7732-18-5	

Test	Specification	Result
Appearance	Clear, colorless	N/A
Water (H <sub>2</sub> O)	0.095-0.105 mg/g	N/A

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
RK410000-10x5R	10x5 mL Glass Ampule	60 months

**Recommended Storage:** 15°C - 30°C (59°F - 86°F)

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.