

## **Product Specification**

## Chromium ICP Standard, 100 ppm Cr in dilute HNO<sub>3</sub>

Lot Number: SAMPLE

Product Number: RPCR01KN

Manufacture Date: N/A

Expiration Date: N/A

Name	CAS#	Grade	
Water	7732-18-5	7732-18-5 ACS/ASTM/USP/EP	
Nitric Acid	7697-37-2	Trace Metals	
Chromium Nitrate Nonahydrate	7789-02-8	High Purity	
Test	Specification	Result	NIST SRM#

Appearance	Colorless to light blue liquid	N/A	
Chromium (Cr)	90-110 ppm	N/A	3112

Reference EPA (200.7)

## Specification

Chromium ICP, 1,000 ppm in HNO3

This standard is guaranteed to be stable and accurate provided the product is kept tightly capped and stored under normal laboratory conditions. Balances are calibrated using NIST traceable weights whose verification of maintenance and recalibration is documented per in-house Standard Operating Procedures. Class A glassware is also calibrated and routinely rechecked per in-house Standard Operating Procedures. Trace metal analyzed acids and Trace Metals Analyzed Water are used in the manufacture of this product. Triple cleaned containers are used in the manufacture of this product.

	Part Number	Size / Package Type	Shelf Life (Unopened Container)
-	RPCR01KN-100N	100 mL natural LDPE	18 months

**Recommended Storage**: 15°C - 30°C (59°F - 86°F)

This document is designed to comply with ISO Guide 31 "Reference Materials --Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.