

Certificate of Analysis

Acetate Buffer, pH 4.0, for Residual Chlorine Analysis

Lot Number: 1307J56

Product Number: 50

Manufacture Date: JUL 11, 2023

Expiration Date: JUN 2025

| Name | CAS# | Grade |
|------------------|-----------|-----------------|
| Acetic Acid | 64-19-7 | ACS |
| Water | 7732-18-5 | ACS/ASTM/USP/EP |
| Sodium Hydroxide | 1310-73-2 | Reagent |

| Test | Specification | Result | NIST SRM# |
|---------------------------------------|---------------------|---------------|-------------------------------|
| Appearance | Colorless liquid | Passed | |
| pH at 25°C (Method: SQCP027, SQCP033) | 3.9-4.1 | 4.1 | 185i, 186-I-g, 186-II-g |
| Specific Gravity | 1.109-1.115 at 25°C | 1.114 at 25°C | |

| Specification | Reference |
|--------------------------------|--------------------|
| Acetate Buffer Solution, pH 4 | APHA (4500-C102 C) |
| Acetate Buffer Solution, pH 4 | APHA (4500-C1 E) |
| Acetate Buffer Solution, pH 4 | APHA (4500-C1 D) |
| Acetate Buffer Solution | APHA (4500-C1 C) |
| Acetate Buffer Solution (pH 4) | EPA (330.2) |
| Acetate Buffer Solution (pH 4) | EPA (330.1) |
| pH 4.0 Buffer Solution | ASTM (D 1253) |

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

| Part Number | Size / Package Type | Shelf Life (Unopened Container) |
|-------------|---------------------|---------------------------------|
| 50-1 | 4 L natural poly | 24 months |
| 50-16 | 500 mL natural poly | 24 months |
| 50-32 | 1 L natural poly | 24 months |

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Heidi J Green (07/11/2023)

Operations Manager

This document is designed to comply with ISO Guide 31 "Reference Materials --
Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.