

# Certificate of Analysis

## Alkaline-Iodide-Azide, Alsterberg Formulation for Dissolved Oxygen (DO) Analysis

**Lot Number:** 1312A64

**Product Number:** 540

**Manufacture Date:** DEC 08, 2023

**Expiration Date:** NOV 2025

This solution is intended for use with samples with normal (saturated or subsaturated) Dissolved Oxygen content.

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Sodium Hydroxide	1310-73-2	Reagent
Potassium Iodide	7681-11-0	ACS
Sodium Azide	26628-22-8	Reagent

Test	Specification	Result
Appearance	Colorless liquid	Passed
Free Iodine	To Pass Test	Passed

Specification	Reference
Alkaline Iodide-Sodium Azide Solution I	ASTM (D 888)
Alkali-Iodide-Azide Reagent, for saturated or less-than-saturated samples	APHA (4500-O D)
Alkali-Iodide-Azide Reagent, for saturated or less-than-saturated samples	APHA (4500-O E)
Alkali-Iodide-Azide Reagent, for saturated or less-than-saturated samples	APHA (4500-O F)
Alkali-Iodide-Azide Reagent, for saturated or less-than-saturated samples	APHA (4500-O C)
Alkaline Iodide-Azide Solution	EPA (360.2)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
540-16	500 mL natural poly	24 months
540-32	1 L natural poly	24 months

**Recommended Storage:** 15°C - 30°C (59°F - 86°F)



Heidi J Green (12/08/2023)

Operations Manager

This document is designed to comply with ISO Guide 31 "Reference Materials --  
Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.