

Certificate of Analysis

Aerosol OT, 10% (w/v)

Lot Number: 1404D20

Product Number: 281

Manufacture Date: APR 01, 2024

Expiration Date: SEP 2025

Reagent Alcohol consists of approximately 90% Ethyl Alcohol (CH₃CH₂OH, CAS No. 64-17-5), 5% Methyl Alcohol (CH₃OH, CAS No. 67-56-1), and 5% Isopropyl Alcohol (CH₃CHOHCH₃, CAS No. 67-63-0).

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Ethyl Alcohol	64-17-5	ACS
Docosate Sodium	577-11-7	Purified
Isopropyl Alcohol	67-63-0	ACS
Methyl Alcohol	67-56-1	ACS

Test	Specification	Result
------	---------------	--------

Appearance Colorless to slightly yellow liquid Passed

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
281-16	500 mL natural poly	18 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Heidi J Green (04/01/2024)

Operations Manager

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.