448 West Fork Dr Arlington, TX 76012 http://www.riccachemical.com 1-888-GO-RICCA

customerservice@riccachemical.com

Certificate of Analysis

Ferrous Iron Standard, 1000 ppm Fe (II)

Lot Number: 1404D40 Product Number: 4198

Manufacture Date: APR 02, 2024

Expiration Date: MAR 2025

The certified value reported is the prepared value based upon the method of preparation of the material. The uncertainty in the prepared value is based upon the volumetric method of preparation.

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Ferrous Ammonium Sulfate Hexahydrate	7783-85-9	ACS
Sulfuric Acid	7664-93-9	ACS

Test	Specification	Result
Appearance	Colorless liquid	Passed
Ferrous Iron (Fe [→])	980-1020 ppm	1000 ppm

Specification	Reference
Ferrous Iron stock solution (1 mg of Fe2+ in 1 mL	ACS (N/A)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
4198-4	120 mL amber glass	12 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

Heidi J Green (04/02/2024)

Operations Manager

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.

Version: 1.3 Lot Number: 1404D40 Product Number: 4198 Page 1 of 1