

Certificate of Analysis

Ammonia Standard, 100 ppm NH₃ (82.2 ppm N)

Lot Number: 1404J37

Product Number: 613

Manufacture Date: APR 19, 2024

Expiration Date: APR 2025

The certified value reported is the prepared value based upon the method of preparation of the material. The uncertainty in the prepared value is based upon the volumetric method of preparation.

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Ammonium Chloride	12125-02-9	High Purity

Test	Specification	Result
Ammonia (NH ₃)	99.5-100.5 ppm	100.0 ppm
Appearance	Colorless liquid	Passed

Specification	Reference
---------------	-----------

Ammonia, Solution, Intermediate (100 mg NH₃ as N/L)

ASTM (D 1426 B)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
613-16	500 mL natural poly	12 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Heidi J Green (04/19/2024)

Operations Manager

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.