

## **Certificate of Analysis**

## Alkaline-Iodide-Azide, Alsterberg Formulation for Dissolved Oxygen (DO) Analysis

## Manufacture Date: APR 19, 2024

Lot Number: 1404J45

Product Number: 540

Expiration Date: APR 2026

This solution is intended for use with samples with normal (saturated or subsaturated) Dissolved Oxygen content.

Name	CAS#	Grade	
Water	7732-18-5	ACS/ASTM/USP/EP	
Sodium Azide	26628-22-8	Reagent	
Potassium Iodide	7681-11-0	ACS	
Sodium Hydroxide	1310-73-2	Reagent	
Test	Specification	Result	

1000	Specification	HOBUIT	
Appearance	Colorless liquid	Passed	
Free Iodine	To Pass Test	Passed	

Specification	Reference
Alkaline Iodide-Sodium Azide Solution I	ASTM (D 888)
Alkali-Iodide-Azide Reagent, for saturated or less-than-saturated samples	APHA (4500-O D)
Alkali-Iodide-Azide Reagent, for saturated or less-than-saturated samples	АРНА (4500-О Е)
Alkali-Iodide-Azide Reagent, for saturated or less-than-saturated samples	АРНА (4500-О F)
Alkali-Iodide-Azide Reagent, for saturated or less-than-saturated samples	АРНА (4500-О С)
Alkaline Iodide Azide Solution	EPA (360.2)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
540-1	4 L natural poly	24 months
540-16	500 mL natural poly	24 months
540-32	1 L natural poly	24 months
D 110		

Recommended Storage: 15°C - 30°C (59°F - 86°F)

Al

Heidi J Green (04/19/2024) Operations Manager

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.