RICCA CHEMICAL COMPANY® NOT THE Pocomoke City, MD 21851 http://www.riccachemical.com 1-888-GO-RICCA customerservice@riccachemical.com

Certificate of Analysis

Buffer, Reference Standard, pH 6.86 ± 0.01 at 25° C In taped cases (no staples) Manufacture Date: SEP 27, 2022 Lot Number: 2209H95 Product Number: 1539 Expiration Date: SEP 2024 The certified value for this product is confirmed in independent testing by a second qualified chemist. °C 0 202530 35 40 505 10 15456.879 pН 6.950 6.9246.899 6.860 6.8526.8446.8406.837 6.836 6.984 Name CAS# Grade Water 7732-18-5 ACS/ASTM/USP/EP Sodium Phosphate Dibasic 7558-79-4 ACS Potassium Dihydrogen Phosphate 7778-77-0 ACS Preservative Proprietary Test Specification Result Passed *Not a certified value. Colorless liquid Appearance Test **Certified Value** Uncertainty NIST SRM# 6.867 0.02pH at 25°C (Method: SQCP027, SQCP033) 185i, 186-I-g, 186-II-g pH measurements were performed in our Pocomoke City, MD laboratory under ISO/IEC 17025 accreditation (ANAB Certificate L2387.01) and are

pH measurements were performed in our Pocomoke City, MD laboratory under ISO/IEC 17025 accreditation (ANAB Certificate L2387.01) and are certified traceable to National Institute of Standards and Technology (NIST) Standard Reference Material as indicated above via an unbroken chain of comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty in the NIST Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in a normal distribution. Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
1539-1	4 L natural poly	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

Muurkey

Myrlande Gilles (09/27/2022) Quality Control This document is designed to comply with ISO Guide 31 "Reference Materials --Contents of Certificates and Labels."

This product was tested in an ISO 17025 Accredited Laboratory

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.

