

Certificate of Analysis

Arsenite Standard, 0.100 Normal (N/10)

Lot Number: 2310P19

Product Number: 820

Manufacture Date: OCT 30, 2023

Expiration Date: OCT 2025

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Arsenic Trioxide	1327-53-3	ACS
Sodium Bicarbonate	144-55-8	ACS
Sodium Hydroxide	1310-73-2	ACS

Test	Specification	Result	NIST SRM#
Appearance	Colorless liquid	Passed	
Assay (vs. Iodine/Starch)	0.0995-0.1005 N at 20°C	0.1004 N at 20°C	136

Specification	Reference
Standard Arsenite Solution, 0.1 N	APHA (4500-C1 C)
Sodium Arsenite (0.1 N)	EPA (SW-846) (9010B)
Sodium Arsenite (0.1 N)	EPA (SW-846) (9012A)
Sodium Arsenite (0.1 N)	EPA (SW-846) (9010B)
Sodium Arsenite (0.1 N)	EPA (SW-846) (9012A)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
820-32	1 L natural poly	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Sharon Travers (10/30/2023)

Operations Manager

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.