

Certificate of Analysis

Barium Chloride, 10% (w/v)
Lot Number: 2402D06

Product Number: 850

Manufacture Date: FEB 12, 2024

Expiration Date: JAN 2026

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Barium Chloride Dihydrate	10326-27-9	ACS

Test	Specification	Result	NIST SRM#
Appearance	Colorless liquid	Passed	
Assay (by Refractive Index)	9.6-10.4 % (w/v)	10.2 % (w/v)	
pH at 25°C (Method: SQCP027, SQCP033)	6.5-6.9	6.7	186-I-g, 186-II-g, 191d

Specification	Reference
Barium Chloride Solution	APHA (4500-SO42- D)
Barium chloride, 10%	TAPPI (T 624 cm-85)
Barium Chloride Solution	APHA (4500-SO42- C)
Barium chloride, 10% solution	TAPPI (T 421 om-97)
Barium Chloride Solution	EPA (375.3)
Barium Chloride Solution (100 g/L)	ASTM (D 1425)
Barium Chloride Solution (100 g/L)	ASTM (D 1266)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
850-1	4 L natural poly	24 months
850-16	500 mL natural poly	24 months
850-2.5	10 L Cubitainer®	24 months
850-32	1 L natural poly	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Jose Pena (02/12/2024)

Operations Manager

This document is designed to comply with ISO Guide 31 "Reference Materials --
Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.