RICCA CHEMICAL COMP.								AN	http://www.riccachemical.com 1-888-GO-RICCA		
customerservice@riccachemical.com											
Buffer, Reference Standard, pH 5.00 ± 0.01 at 25° C											
Lot Number: 2404E07 Pro				roduct	oduct Number: 1505				Manufacture Date: APR 08, 2024 Expiration Date: MAR 2026		
The certified value for this product is confirmed in independent testing by a second qualified chemist. The NIST Traceable pH value is certified to ±0.01 at 25 °C only. All other pH values at their corresponding temperatures are accurate to ± 0.05.											
°C pH	$\begin{array}{c} 10 \\ 4.96 \end{array}$	$\begin{array}{c} 15 \\ 4.97 \end{array}$	$\begin{array}{c} 20 \\ 4.98 \end{array}$	$\begin{array}{c} 25\\ 5.00 \end{array}$	$\begin{array}{c} 30 \\ 5.01 \end{array}$	$\begin{array}{c} 35\\ 5.03\end{array}$	$\begin{array}{c} 40 \\ 5.04 \end{array}$	$\begin{array}{c} 45\\ 5.06\end{array}$	$\begin{array}{c} 50 \\ 5.08 \end{array}$		
Name					CAS#				Grade		
Water					7732-18-5				ACS/ASTM/USP/EP		
Potassium Acid Phthalate						877-24-7				Buffer	
Sodium Hydroxide						1310-73-2				Reagent	

PreservativeProprietaryTestSpecificationResultAppearanceColorless liquidPassed*Not a certified value.TestCertified ValueUncertaintyNIST SRM#pH at 25°C (Method: SQCP027_SQCP033)4.9940.02185i 186-I-g 186-II-g

pH at 25°C (Method: SQCP027, SQCP033)4.9940.02185i, 186-I-g, 186-II-gpH measurements were performed in our Pocomoke City, MD laboratory under ISO/IEC 17025 accreditation (ANAB Certificate L2387.01) and are
certified traceable to National Institute of Standards and Technology (NIST) Standard Reference Material as indicated above via an unbroken chain of
comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty in the NIST
Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in
a normal distribution. Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated
before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with
weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated
regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture
according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
1505-5	20 L Cubitainer®	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

Jose Pena (04/08/2024) Operations Manager This document is designed to comply with ISO Guide 31 "Reference Materials --Contents of Certificates and Labels."

This product was tested in an ISO 17025 Accredited Laboratory

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.