RICCA CHEMICAL COMPANY® 1841 Broad Street Pocomoke City, MD 21851 http://www.riccachemical.com 1-888-GO-RICCA customerservice@riccachemical.com

Certificate of Analysis

Buffer, Reference Standard, pH 6.86 ± 0.01 at 25° C In taped cases (no staples)

Lot Number: 2404G46				Product Number: 1539								Cacture Date: APR 17, 2024 Expiration Date: APR 2026	
The certified value for this product is confirmed in independent testing by a second qualified chemist. C = 0 5 10 15 20 25 20 45 50													
°C pH ($\begin{array}{c} 0\\ 5.984\end{array}$	$\begin{array}{c} 5 \\ 6.950 \end{array}$	$\begin{array}{c} 10 \\ 6.924 \end{array}$	$\begin{array}{c} 15 \\ 6.899 \end{array}$	$\begin{array}{c} 20 \\ 6.879 \end{array}$	$\begin{array}{c} 25 \\ 6.860 \end{array}$	$\begin{array}{c} 30 \\ 6.852 \end{array}$	$\begin{array}{c} 35 \\ 6.844 \end{array}$	$\begin{array}{c} 40 \\ 6.840 \end{array}$	$\begin{array}{c} 45 \\ 6.837 \end{array}$	$\begin{array}{c} 50 \\ 6.836 \end{array}$		
Name						CAS#				Grade			
Water						7732-18-5				ACS/A	STM/USP/	ΈP	
Sodium Phosphate Dibasic						7558-79-4				ACS			
Potassium Dihydrogen Phosphate						7778-77-0				ACS			
Preserv	ative					Pr	oprietai	у					
Test						Specification				R	esult		
Appearance						Colorless liquid				Passed		*Not a certified value	
Test						Certified Value				Uncertainty		NIST SRM#	
pH at 25°C (Method: SQCP027, SQCP033)							6.866				02	185i, 186-I-g, 186-II-g	
pH measurements were performed in our Pocomoke City, MD laboratory under ISO/IEC 17025 accreditation (ANAB Certificate L2387.01) and are certified traceable to National Institute of Standards and Technology (NIST) Standard Reference Material as indicated above via an unbroken chain of comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty in the NIST Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in a normal distribution. Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.													
Part Number Siz						ze / Package Type					Shelf Life (Unopened Container)		
											0.4 1		

1539-1 4 L natural poly

24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

Jose Pena (04/17/2024) Operations Manager This document is designed to comply with ISO Guide 31 "Reference Materials --Contents of Certificates and Labels."

This product was tested in an ISO 17025 Accredited Laboratory

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.