

Certificate of Analysis

Ammonium Chloride-EDTA Solution for Nitrate by the Cadmium Reduction Method

Lot Number: 2404H05

Product Number: 626

Manufacture Date: APR 19, 2024

Expiration Date: APR 2025

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Ammonium Chloride	12125-02-9	ACS
Ethylenediaminetetraacetic Acid (EDTA), Disodium Salt, Dihydrate	6381-92-6	ACS
Ammonium Hydroxide	1336-21-6	ACS

Test	Specification	Result	NIST SRM#
Appearance	Colorless liquid	Passed	
pH at 25°C (Method: SQCP027, SQCP033)	8.48-8.52	8.50	186-I-g, 186-II-g, 191d

Specification	Reference
Ammonium Chloride-EDTA Solution	APHA (4500-NO3- E)
Ammonium Chloride-EDTA Solution	EPA (353.3)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
626-1	4 L natural poly	12 months
626-16	500 mL natural poly	12 months
626-2.5	10 L Cubitainer®	12 months
626-32	1 L natural poly	12 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Jose Pena (04/19/2024)
Operations Manager

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.