

# Certificate of Analysis

**Ammonium Phosphate Solution, 400 g (NH<sub>4</sub>)<sub>2</sub>HPO<sub>4</sub> + 800 mL Water**

**Lot Number:** 4311G40

**Product Number:** 690

**Manufacture Date:** NOV 13, 2023

**Expiration Date:** OCT 2025

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Ammonium Phosphate Dibasic	7783-28-0	ACS

Test	Specification	Result
Appearance	Colorless liquid	Passed

Specification	Reference
Diammonium Hydrogen Phosphate Solution	APHA (4500-S2- D)
Diammonium Hydrogen Phosphate Solution	EPA (376.2)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
690-1	4 L natural poly	24 months

**Recommended Storage:** 15°C - 30°C (59°F - 86°F)



Paul Brandon (11/13/2023)

Production Manager

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.