

Certificate of Analysis

Acetic Acid, 5% (v/v)**Lot Number:** 4312J07**Product Number:** 130**Manufacture Date:** DEC 21, 2023**Expiration Date:** DEC 2026

| Name | CAS# | Grade |
|-------------|-----------|-----------------|
| Water | 7732-18-5 | ACS/ASTM/USP/EP |
| Acetic Acid | 64-19-7 | ACS |

| Test | Specification | Result | NIST SRM# |
|--|-------------------|--------------|-----------|
| Appearance | Colorless liquid | Passed | |
| Assay (vs. Sodium Hydroxide/Thymol Blue) | 4.94-5.06 % (v/v) | 5.02 % (v/v) | 84 |

| Specification | Reference |
|-----------------------------|---------------|
| Acetic Acid Solution | ASTM (D 4045) |
| Acetic Acid Solution (1+19) | ASTM (E 200) |

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

| Part Number | Size / Package Type | Shelf Life (Unopened Container) |
|-------------|---------------------|---------------------------------|
| 130-1 | 4 L natural poly | 36 months |
| 130-16 | 500 mL natural poly | 36 months |
| 130-32 | 1 L natural poly | 36 months |

Recommended Storage: 15°C - 30°C (59°F - 86°F)

Paul Brandon (12/21/2023)

Production Manager

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.