

Certificate of Analysis

Aluminum AA Standard, 1000 ppm Al in 3% HNO₃

Lot Number: 4401S58

Product Number: AAL1KN

Manufacture Date: JAN 26, 2024

Expiration Date: JAN 2026

The certified value for this product is confirmed in independent testing by a second qualified chemist.
The final solution concentration is confirmed by AA, ICP, or ICP-MS, and is traceable to NIST Standard Reference Material 3101a.

Name	CAS#	Grade
Water	7732-18-5	
Nitric Acid	7697-37-2	
Aluminum Nitrate Nonahydrate	7784-27-2	

Test	Specification	Result	NIST SRM#
Appearance	Colorless liquid	Passed	
Assay (vs. EDTA/Dithizone)	995-1005 ppm Al	1000 ppm Al	915

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
AAL1KN-100	100 mL natural LDPE	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Paul Brandon (01/26/2024)

Production Manager

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.