RICCA CHEMICAL COMPANY®

Arlington, TX 76012
Pocomoke City, MD 21851
Batesville, IN 47006
http://www.riccachemical.com
1-888-GO-RICCA
customerservice@riccachemical.com

Product Specification

Reference Color Standard No. 6, ASTM D 848

Lot Number: SAMPLE **Product Number:** R6605600

Manufacture Date: N/A
Expiration Date: N/A

This solution is prepared volumetrically based on specifications found in ASTM D 848-02.

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Ferric Chloride Hexahydrate	10025-77-1	ACS
Hydrochloric Acid	7647-01-0	ACS
Cobalt (II) Chloride Hexahydrate	7791-13-1	ACS

Test	Specification	Result
Appearance	Light yellow-brown liquid	N/A

Specification	Reference
Reference Color Standard, No. 6	ASTM (D 848-02)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
R6605600-100C	100 mL amber glass	12 months
R6605600-500C	500 mL amber glass	12 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.