

Product Specification

Calcium Standard, 10 ppm Ca

Lot Number: SAMPLE**Product Number:** 1790**Manufacture Date:** N/A**Expiration Date:** N/A

The certified value reported is the prepared value based upon the method of preparation of the material. The uncertainty in the prepared value is based upon the volumetric method of preparation.

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Nitric Acid	7697-37-2	Trace Metals
Calcium Carbonate	471-34-1	ACS Chelometric

Test	Specification	Result	NIST SRM#
Appearance	Colorless liquid	N/A	
Calcium (Ca)	9.92-10.08 ppm	N/A	3109

Specification	Reference
Calcium Standard, 1 mL = 10 µg Ca, 10 ppm Ca	EPA (215.1)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
1790-16	500 mL natural poly	12 months
1790-4	120 mL natural poly	12 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.