

# Product Specification

## Clarke Solution

Lot Number: SAMPLE

Product Number: R2137000

Manufacture Date: N/A

Expiration Date: N/A

Name	CAS#	Grade
Water	7732-18-5	ACS
Hydrochloric Acid	7647-01-0	ACS
Stannous Chloride Dihydrate	10025-69-1	ACS
Antimony Trioxide	1309-64-4	Reagent

Test	Specification	Result
------	---------------	--------

Appearance	Clear, tan liquid	N/A
------------	-------------------	-----

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
R2137000-1C	1 L amber glass	12 months

**Recommended Storage:** 15°C - 30°C (59°F - 86°F)

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.