

Product Specification

Diphenylcarbazone-Bromophenol Blue Mixed Indicator, for Chloride Determination

Lot Number: SAMPLE

Product Number: 2610

Manufacture Date: N/A

Expiration Date: N/A

Name	CAS#	Grade
Ethyl Alcohol	64-17-5	Reagent
Methyl Alcohol	67-56-1	Reagent
s-Diphenylcarbazone	538-62-5	ACS
Diphenylcarbazide	140-22-7	ACS
Bromophenol Blue, Sodium Salt	62625-28-9	ACS

Test	Specification	Result
Appearance	Deep red liquid	N/A
Suitability for Use	Yellow (excess Chloride & proper pH) - Purple (excess Mercury)	N/A

Specification	Reference
Mixed Indicator Solution	ASTM (D 512 A)
Mixed Indicator Reagent	APHA (4500-Cl- C)
Mixed Indicator Reagent	EPA (325.3)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
2610-1	4 L amber glass	9 months
2610-16	500 mL amber glass	9 months
2610-32	1 L amber glass	9 months
2610-4	120 mL amber glass	9 months

Recommended Storage: 2°C - 8°C (36°F - 46°F)

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.