

Product Specification

Murexide Indicator, 0.1% (w/w) in Ammonium Chloride

Lot Number: SAMPLE

Product Number: R5219000

Manufacture Date: N/A

Expiration Date: N/A

Name	CAS#	Grade
Ammonium Chloride	12125-02-9	ACS
Murexide	3051-09-0	ACS

Test	Specification	Result
Appearance	Rose-pink solid	N/A
Suitability for Use	Pink (Excess Calcium) - Purple (Excess EDTA)	N/A

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
R5219000-100A6	100 g poly	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.