

Product Specification

Nitrite Standard, 1000 ppm NO₂⁻ (304.5 ppm N)

Lot Number: SAMPLE

Product Number: 5445

Manufacture Date: N/A

Expiration Date: N/A

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Sodium Nitrite	7632-00-0	ACS
Chloroform	67-66-3	

Test	Specification	Result	NIST SRM#
Appearance	Colorless liquid	N/A	
Assay (vs. Potassium Permanganate)	996.7-1003.3 ppm NO ₂	N/A	8040

Specification	Reference
Standard Nitrite Solution, 1000 mg/L	APHA (4110 B)
Nitrite Stock Solution (1.00 mL = 1.00 mg NO ₂ ⁻)	EPA (SW-846) (9056)
Nitrite Stock Solution (1.00 mL = 1.00 mg NO ₂ ⁻)	ASTM (D 4327)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
5445-16	500 mL natural poly	12 months
5445-4	120 mL natural poly	12 months
5445-8	250 mL natural poly	12 months

Recommended Storage: 2°C - 8°C (36°F - 46°F)

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.