

Manufacture Date: N/A

## **Product Specification**

## Sodium Hydroxide, 0.0833 Normal

Lot Number: SAMPLE Prod	uct Number: 7348		Expiration Date: N/A	
Name	CAS#	Grade		
Water	7732-18-5	ACS/ASTM/USP/EP		
Sodium Hydroxide	1310-73-2	Reagent		
Test	Specification	Result	NIST SRM#	
Appearance	Colorless liquid	N/A		
Assay (vs. Potassium Hydrogen Phthalate/Phenolphthalein)	0.08325-0.08335 N at	t 20°C N/A	84	
Volumetric glassware complies with Class A toler	ance requirements of ASTM E 288 and N	IST Circular 434; it is calibrat	ed before first use and	

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
7348-5	20 L Cubitainer®	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

This document is designed to comply with ISO Guide 31 "Reference Materials --Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.