

# Product Specification

## Mixed AA Standard

**Lot Number:** SAMPLE

**Product Number:** AAMIX02

**Manufacture Date:** N/A

**Expiration Date:** N/A

This is a multielement solution that was prepared volumetrically to contain the elements and concentrations stated below. The uncertainty associated with the certified values is  $\pm 0.5\%$  relative. The concentrations are confirmed by ICP or ICP-MS.

**Matrix:** 5% Nitric Acid

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Nitric Acid	7697-37-2	Trace Metals

Test	Specification	Result
Appearance	Colorless liquid	N/A

Analyte	Analysis (ppm)	Solute	CAS#	Grade	NIST SRM#
Potassium (K)	995-1005 ppm	Potassium Nitrate	7757-79-1	High Purity	3141
Calcium (Ca)	995-1005 ppm	Calcium Carbonate	471-34-1	High Purity	3109
Sodium (Na)	995-1005 ppm	Sodium Carbonate	497-19-8	High Purity	3152
Magnesium (Mg)	995-1005 ppm	Magnesium	7439-95-4	High Purity	3131

This standard is guaranteed to be stable and accurate provided the product is kept tightly capped and stored under normal laboratory conditions. Balances are calibrated using NIST traceable weights whose verification of maintenance and recalibration is documented per in-house Standard Operating Procedures. Class A glassware is also calibrated and routinely rechecked per in-house Standard Operating Procedures. Trace metal analyzed acids and Trace Metals Analyzed Water are used in the manufacture of this product. Triple cleaned containers are used in the manufacture of this product.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
AAMIX02-500	500 mL natural poly	12 months

**Recommended Storage:** 15°C - 30°C (59°F - 86°F)

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.