

Product Specification

ProteoSpec[™] MSTFA
N-Methyl-N-(trimethylsilyl)trifluoroacetamide
Lot Number: SAMPLE

Product Number: RMB12104

Date Packaged: N/A

Expiration Date: N/A

Contains
 $C_6H_{12}F_3NOSi$ CAS No. MW 199.25

Test	Specification	Result
------	---------------	--------

Appearance	Clear, colorless liquid	N/A
------------	-------------------------	-----

Assay (by GC)	min 98.0 %	N/A
---------------	------------	-----

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
-------------	---------------------	---------------------------------

RMB12104-10x1C2	10 x 1g glass case	12 months
-----------------	--------------------	-----------

RMB12104-10x1R	10 x 1mL case	12 months
----------------	---------------	-----------

Recommended Storage: 15°C - 30°C (59°F - 86°F)

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.