

Certificate of Analysis

Potassium Hydroxide, 50% (w/v), Suitable for Orsat Gas Analysis

Lot Number: 1110370

Product Number: 6160

Manufacture Date: OCT 01, 2021

Expiration Date: SEP 2024

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Potassium Hydroxide	1310-58-3	Reagent

Test	Specification	Result	NIST SRM#
Appearance	Colorless liquid	Passed	
Assay (vs. Sulfuric Acid/Phenolphthalein)	49.94-50.06 % (w/v)	50.03 % (w/v)	723

Specification	Reference
Potassium hydroxide solution	APHA (2720 B)
Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.	

Part Number	Size / Package Type	Shelf Life (Unopened Container)
6160-1	4 L natural poly	36 months
6160-16	500 mL natural poly	36 months
6160-5	20 L Cubitainer®	36 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Israel Alamudun (10/01/2021)

Quality Control Supervisor

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.