

Certificate of Analysis

Nitrate Standard, 10,000 ppm NO₃⁻

Lot Number: 1306M48

Product Number: R5307990

Manufacture Date: JUN 27, 2023

Expiration Date: JUN 2024

The certified value reported is the prepared value based upon the method of preparation of the material. The uncertainty in the prepared value is based upon the volumetric method of preparation.

| Name | CAS# | Grade |
|-------------------|-----------|-----------------|
| Water | 7732-18-5 | ACS/ASTM/USP/EP |
| Potassium Nitrate | 7757-79-1 | High Purity |

| Test | Specification | Result |
|----------------------------|------------------|-----------|
| Appearance | Colorless liquid | Passed |
| Nitrate (NO ₃) | 9950-10050 ppm | 10000 ppm |

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

| Part Number | Size / Package Type | Shelf Life (Unopened Container) |
|-------------|---------------------|---------------------------------|
| R5307990-1A | 1 L natural poly | 12 months |

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Heidi J Green (06/27/2023)

Operations Manager

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.