

## **Certificate of Analysis**

## Nitrate Standard, 1000 ppm NO<sub>3</sub> (226 ppm N)

## Lot Number: 1403K05

Product Number: 5307

## Manufacture Date: MAR 22, 2024 Expiration Date: SEP 2025

The certified value reported is the prepared value based upon the method of preparation of the material. The uncertainty in the prepared value is based upon the volumetric method of preparation.

CAS#	Grade
7732-18-5	ACS/ASTM/USP/EP
7757-79-1	High Purity
Specification	Result
Colorless liquid	Passed
995-1005 ppm	1000 ppm
	Reference
	EPA (SW-846) (9210)
	EPA (SW-846) (9056)
	ASTM (D 5996)
	ASTM (D 4327)
	7732-18-5 7757-79-1 <b>Specification</b> Colorless liquid 995-1005 ppm

recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
5307-16	500 mL natural poly	18 months
<b>D</b> ecommonded Store as: $15\%$ - $20\%$ ( $50\%$ - $8\%$ )		

Recommended Storage: 15°C - 30°C (59°F - 86°F)

Heidi J Green (03/22/2024) Operations Manager This document is designed to comply with ISO Guide 31 "Reference Materials --Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.