

Certificate of Analysis

Nitrite Nitrogen Standard, 1000 ppm N (3285 ppm NO₂)

Lot Number: 1404G39

Product Number: 5461

Manufacture Date: APR 15, 2024

Expiration Date: OCT 2024

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Chloroform	67-66-3	
Potassium Nitrite	7758-09-0	ACS

Test	Specification	Result	NIST SRM#
Appearance	Colorless liquid	Passed	
Assay (vs. Potassium Permanganate)	995-1005 ppm N	999 ppm N	8040

Specification	Reference
Nitrite Solution, Stock (1.0 mL = 1.0 mg NO ₂ -N)	ASTM (D 3867 A)
Nitrite Solution, Stock (1.0 mL = 1.0 mg NO ₂ -N)	ASTM (D 3867 B)
Stock Nitrite Solution: 1 mL = 1.0 mg NO ₂ -N	EPA (353.2)
Stock Nitrite Solution: 1.0 mL = 1.00 mg NO ₂ -N	EPA (353.3)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
5461-4	120 mL natural poly	6 months

Recommended Storage: 2°C - 8°C (36°F - 46°F)



Heidi J Green (04/15/2024)

Operations Manager

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.