

# Certificate of Analysis

## Turbidity Standard, 8 NTU, Polymer Bead Suspension

**Lot Number:** 1505K15**Product Number:** R8801800**Manufacture Date:** MAY 10, 2025**Expiration Date:** MAY 2026

Ready to use. Do NOT agitate. Pour gently into cuvette avoiding bubbles which will interfere with Turbidity readings.

There is no NIST Standard Reference Material for Turbidity. This solution is traceable to commercially prepared EPA recognized Primary Standards.

The EPA primary standard for turbidity (Formazin suspension, 4000 NTU) is identical to EP Primary Opalescent Suspension and USP Formazin Stock Suspension.

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Preservative	Proprietary	
Polystyrene Microspheres	9003-53-6	Reagent

Test	Specification	Result
Appearance	Colorless liquid	Passed
Turbidity	7.9-8.1 NTU	8.0 NTU

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
R8801800-100C	100 mL amber glass	12 months

**Recommended Storage:** 15°C - 30°C (59°F - 86°F)

Ernest Mahan (05/10/2025)

Plant Manager

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.