

Certificate of Analysis

Cobalt Standard, 10,000 ppm Co

Lot Number: 2401D06 Produ	act Number: 2205		oate: JAN 08, 202 on Date: JUN 202
Name	CAS#	Grade	
Water	7732-18-5	ACS/ASTM/USP/EP	
Cobalt Nitrate Hexahydrate	10026-22-9	ACS	
Nitric Acid	7697-37-2	ACS	
Test	Specification	Result	NIST SRM#
Appearance	Pink liquid	Passed	
Assay (vs. EDTA/Xylenol Orange)	9990-10010 ppm Co	10007 ppm Co	915
Specification	Re	ference	
Cobalt Standard, 1 mL = 10 mg Co, 10,00	0 ppm Co EP	A (219.1)	
Volumetric glassware complies with Class A toleral recalibrated regularly in accordance with ASTM E traceable to the NIST national mass standard. The thermometer traceable to NIST standards. All pro- methods. Batch records document raw material tra	542 and NIST Procedure NBSIR 74-461. rmometers and temperature probes are clucts are prepared according to master do	Balances are calibrated regularly we alibrated before first use and recalibrated before that assure manufacture a	vith weights certified ibrated regularly with a
Part Number	Size / Package Type	Shelf Life (Unoper	ned Container)
2205-16	500 mL natural poly	18 months	

Recommended Storage: 15°C - 30°C (59°F - 86°F)

Jose Pena (01/08/2024) Operations Manager This document is designed to comply with ISO Guide 31 "Reference Materials --Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.