

Certificate of Analysis

Hexavalent Chromium Standard, 100 ppm Cr⁶⁺

Lot Number: 2403J26

Product Number: 2095

Manufacture Date: MAR 18, 2024

Expiration Date: SEP 2025

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Potassium Dichromate	7778-50-9	ACS

Test	Specification	Result	NIST SRM#
Appearance	Orange liquid	Passed	
Assay (vs. Sodium Thiosulfate/Starch)	99.5-100.5 ppm Cr	99.9 ppm Cr	136

Specification	Reference
Cr(VI) stock solution, 100 mg/L Cr ⁶⁺	APHA (3500-Cr C)
Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.	

Part Number	Size / Package Type	Shelf Life (Unopened Container)
2095-4	120 mL natural poly	18 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

Jose Pena (03/18/2024)

Operations Manager

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.