## RICCA CHEMICAL COMPANY®

## **Certificate of Analysis**

## Hydrochloric Acid, 10% (v/v)

Lot Number: 2404H61 Product Nu	mber: 3570		Manufacture Date: APR 22, 202 Expiration Date: APR 202	
Name	CAS#	Grade		
Water	7732-18-5	ACS/ASTM/USP/EP		
Hydrochloric Acid	7647-01-0	ACS		
Test	Specification	Result	NIST SRM#	
Appearance	Colorless liquid	Passed		
Assay (vs. Sodium Hydroxide/Phenolphthalein)	9.92-10.08 % (v/v)	10.08 % (v/v)	84	
Specification	Reference			
Hydrochloric Acid (1+9)	ASTM (D 287 M)			
Hydrochloric acid, 1+9	TAPPI (T 419 om-91)			
Hydrochloric Acid (1+9)	ASTM (D 2036 D)			
Hydrochloric Acid (1+9)	ASTM (D 2187 E)			
Hydrochloric, Acid, 1 + 9	APHA (3500-Mg D)			
Hydrochloric Acid, 1+ 9	APHA (4500-CN- K)			
Hydrochloric Acid (1+9)	ASTM (D 2187 F)			
Hydrochloric Acid (1+9)	ASTM (D 2187 G)			
Hydrochloric Acid (1+9)	ASTM (D 2187 H)			
Hydrochloric Acid (1+9)	ASTM (D 4548)			
Hydrochloric Acid (1+9)	ASTM (D 4994 A)			
Hydrochloric Acid (1+9)	ASTM (D 2187 J)			
Hydrochloric Acid Solution (1+9)	ASTM (D 2187 K)			
Hydrochloric Acid (1+9)	ASTM (D 2187 A)			
Hydrochloric Acid (1+9)	ASTM (D 2187 L)			
$II_{-1} = 1 = 1 = 1 = 1 = 1 = 1 = 1 = 1 = 1$	ASTM (D 2026 A1)			

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)		
3570-1	4 L natural poly	36 months		
3570-32	1 L natural poly	36 months		

Recommended Storage: 15°C - 30°C (59°F - 86°F)

Hydrochloric Acid (1+9)

ASTM (D 2036 A1)

 $\left( \right)$ Jose Pena (04/22/2024)

Jose Pena (04/22/2024) Operations Manager

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.