

Certificate of Analysis

Phosphoric Acid, 30% (v/v)

Lot Number: 4311R32

Product Number: 5851.5

Manufacture Date: NOV 30, 2023

Expiration Date: NOV 2025

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Phosphoric Acid	7664-38-2	ACS

Test	Specification	Result	NIST SRM#
Appearance	Colorless liquid	Passed	
Assay (vs. Sodium Hydroxide/Thymolphthalein)	29.96-30.04 % (v/v)	30.03 % (v/v)	84

Specification	Reference
Phosphoric Acid Solution, 1+9	APHA (5530 B)
Phosphoric acid, 10% (v/v)	APHA (3113 B)
Phosphoric Acid Solution, 1 + 9	APHA (4500-C1 C)
Distillation Reagent	EPA (SW-846) (9066)
Distillation Reagent	EPA (420.2)
Phosphoric Acid Solution, 1+9	EPA (420.1)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
5851.5-32	1 L natural poly	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Paul Brandon (11/30/2023)

Production Manager

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.