

# Certificate of Analysis

## Xylenol Orange, 0.1% in Sodium Chloride

Lot Number: 4403G56

Product Number: R9394000

Manufacture Date: MAR 11, 2024

Expiration Date: FEB 2026

Name	CAS#	Grade
Sodium Chloride	7647-14-5	ACS
Xylenol Orange Sodium Salt	3618-43-7	ACS

Test	Specification	Result
Appearance	Red-brown solid	Passed
Suitability for Use	Violet-Red to Purple (excess Metal) - Yellow (excess EDTA)	Passed

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
R9394000-100C2	100 g amber glass	24 months

**Recommended Storage:** 15°C - 30°C (59°F - 86°F)

Paul Brandon (03/11/2024)

Production Manager

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.