

# Certificate of Analysis

## Phosphate Phosphorus Standard, 50 ppm P (153 ppm PO<sub>4</sub><sup>3-</sup>)

**Lot Number:** 4404E86

**Product Number:** 5830

**Manufacture Date:** MAR 08, 2024

**Expiration Date:** MAR 2025

Use purified water to make dilutions of this standard.

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Potassium Dihydrogen Phosphate	7778-77-0	High Purity

Test	Specification	Result	NIST SRM#
Appearance	Colorless liquid	Passed	
Phosphorus (P)	49.75-50.25 ppm	50.00 ppm	3139

Specification	Reference
Stock Solution (1.0 mL = 0.05 mg Phosphorus)	ASTM (D 515 A)
Standard Phosphate Solution	APHA (4500-P C)
Standard Phosphate Solution	APHA (4500-P D)
Stock Phosphate Solution	APHA (4500-P E)
Stock Phosphorus Solution, 1 mL = 0.05 mg P	EPA (365.2)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
5830-4	120 mL natural poly	12 months

**Recommended Storage:** 15°C - 30°C (59°F - 86°F)



Paul Brandon (03/08/2024)

Production Manager

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.